

# Development of reference instruments for multi-wavelength absorption measurements

***Thomas Müller,***  
***Leibniz Institute for Tropospheric Research, Leipzig***

***ACTRIS-2 WP3 Workshop Athens 10-12 November 2015***



# Introduction

The goal was to setup a reference instrument for measuring the light absorption coefficient of airborne particles



- Reference for Multi-Wavelength Absorption photometers
- Calibration of reference instrument should be linked to fundamental physical measurements (SI traceable)

The method of choice is “*extinction minus scattering*”


Reference setup consist of commercial available instruments

- One three wavelength Nephelometer (Aurora 4000) for measuring the scattering coefficient
- Three Cavity Attenuated Phase Shift (CAPS) monitors for measuring the extinction coefficient

# Working principle



Instruments to be tested



## Factors affecting the uncertainty of the reference absorption




- ① effective cell length
- ② uncertainty of CO<sub>2</sub> calibration
- ③ truncation correction
- ④ Wavelength interpolation
- ⑤ Particle transport losses
- ⑥ Adjustment factor
- ⑦ Zero drift

## Further factor affecting the precision of intercomparison

- a. CAPS noise
- b. Aurora noise
- c. AE33 noise
- d. MAAP noise

# CAPS and Nephelometer tests using ammonium sulfate

Truncation correction scattering vs. extinction coefficients



- Nonlinearity of CAPS must be corrected for
- CAPS is about 7-8% lower than Nephelometer

## 'Calibrations of CAPS'


- Massoli et al. (2010) confirmed CAPS to be precise within 10%.
- Petzold et al. (2012) tested CAPS using PSL & Mie calculations.
  - The recalibration showed that the **extinction is low by 5%**. A factor of 1.05 was included in the firmware of all CAPS.
- Laboratory experiment (Sep to Nov 2015) at TROPOS. Experiments with ammonium sulfate showed that **extinction is about 6-7% higher** than scattering measured by Nephelometer.

## Possible reason for discrepancy

- Petzold et al (2012) showed Mie scattering calculations with fitted log normal size distributions. The **standard deviation of fitted modes might be too large**, since the sizing instrument (OPC, Grimm 1.129) can not resolve PSL size distributions.
- Recalculation of data (see Tab. 2 in Petzold et al., 2012), but with typical standard deviations for PSL of 3%
  - **Extinction calculated with fitted size distributions is 5% higher than with a narrow PSL size distribution.**

# Uncertainty of the Reference Absorption System

Scattering  
Nephelometer



## Calibration error

- Purity of calibration gas
- ...
- **Error about 3%**

## Error of truncation correction $\Delta C_{ts}$

- Light source geometry factors
- Aerosol phase function (refractive index, size, etc..  $\Leftrightarrow$  requires iterative processing)
- **Error approx. 5% (Bond et al. 2009)**

Extinction  
CAPS



## Calibration error $\Delta C_{cal}$


- Effective path length error about 7 % ?
- **Can be minimized to about 3% by coupling to Nephelometer.**

Absorption  
=Extinction - Scattering



Total error **approx. 8%**

# Uncertainty of the Reference Absorption System



# Traceability of reference absorption to SI

## **Nephelometer scattering:** factors that need to be considered

- Rayleigh scattering of calibration gas
- Truncation correction can be calculated
  - geometrical factors (cell design)
  - simple polar intensity measurement (light source)
  - Mie scattering calculation for well know particles (e.g. PSL, NIST traceable)

## **CAPS extinction:** factors that need to be considered

- Attenuation of light beam measurement transformed to a phase shifted signal (linearity ?)
- Effective path length can be ‘calibrated’ by adjusting to scattering (nephelometer)

**Does the reference absorption system fulfil the requirements to be ‘SI traceable’ ?**

# Calibration/-checks of the Reference Absorption System

Calibration type	Interval	uncertainty
Zero check of CAPS	15 minutes	0.1 Mm <sup>-1</sup> higher for 'red' CAPS
Zero check of Nephelometer	daily	0.1 Mm <sup>-1</sup>
Nephelometer calibration (CO <sub>2</sub> )	weekly	<3%
Check of CAPS effective pathlength factor	weekly	<3%

On basis of about 6 weeks continuous measurements the reference absorption system was stable within the uncertainty.

More calibration cycles needed for assessing the long term stability.